



228 Business Center Drive Reisterstown, MD 21136 | 1.800.359.0911 x3

Control Summary

Sorted By Control
 Date Range: 4/1/2013 to 4/30/2013
 Test Range: 110 to 290
 Control Range: .C1 to .C1

Control Name	Test	Lot #	# Runs	Expected		Calculated		
				Mean	2SD	Mean	2SD	%CV
Low Chemistry								
	Glucose	CHEM005	30	65	6	65	6	5
	BUN	CHEM005	30	6	2	6	2	17
	Creatinine	CHEM005	30	2.0	1.0	2.0	0.8	19.8
	B/C Ratio	CHEM005	30	0.0	0.0	3.4	2.2	32.9
	Sodium	CHEM005	30	128	3	128	3	1
	Potassium	CHEM005	30	2.8	0.2	2.8	0.2	4.2
	Chloride	CHEM005	30	92	2	92	2	1
	CO2	CHEM005	30	15	3	16	3	9
	Anion Gap	CHEM005	30	0.0	0.0	20.8	5.7	13.7
	Calcium	CHEM005	30	7.8	0.6	7.8	0.5	2.9
	Phosphorus	CHEM005	30	1.9	0.2	1.9	0.2	4.8
	Total Protein	CHEM005	30	3.8	0.8	4.8	11.8	123.0
	Albumin	CHEM005	30	1.7	0.8	1.8	0.7	19.5
	Uric Acid	CHEM005	30	2.1	0.5	2.1	0.4	9.3
	AST	CHEM005	30	12	4	11	4	16
	ALT	CHEM005	30	10	2	10	2	11
	Alk Phos	CHEM005	30	28.0	4.0	28.6	4.2	7.3